TempSpan Transparent and TempSpan CMT

English

Instructions for Use

TempSpan® CMT Temporary Cement
TempSpan® Transparent Temporary Cement

TempSpan Temporary Cements are resin based, eugenol-free cements designed for retaining provisional restorations. Both cements feature our convenient double barrel syringe with auto-mix tips that enable you to place cement directly into the provisional restoration.

Instructions for Use
Indications
- Cementation of temporary crowns, bridges, inlays, and onlays

Contraindications
- Permanent cementation of definitive restorations
- Temporary cementation of provisional veneers

At the ideal mixing conditions of 70-77 °F (21-25 °C) and 40-60% relative humidity, the following work time and setting time from start of mixing are obtained.

Temporary Cement	Work Time	Set Time
TempSpan® CMT Temporary Cement	1 - 1½ minutes	2½ - 3 minutes
TempSpan® Transparent Temporary Cement	≥1½ minutes	≤5½ minutes

1. Prior to placing the mixing tip onto the syringe express cement until both sides flow evenly.

2. Dry the prepared teeth and the internal surface of the restoration.

3. Using slight pressure, dispense a thin layer of TempSpan® Temporary Cement directly into the provisional restoration.

4. Within the specified work time, carefully, but firmly, seat the provisional restoration on the preparation site.

5. Instruct the patient to gently occlude into the provisional until the restoration is fully seated.

6. Clean any excess TempSpan® Temporary Cement with a cotton swab prior to setting.

7. TempSpan® CMT Temporary Cement will set in approximately 2½ - 3 minutes from the start of mix. TempSpan® Transparent Temporary Cement will set in approximately 5½ minutes from the start of mix. TempSpan® Transparent Temporary Cement can also be light cured 20 seconds per surface to accelerate the set of the cement.

8. Remove set cement with a scaler or other suitable instrument.

CAUTION
TempSpan® Temporary Cements contains (meth) acrylic resins. Uncured (meth) acrylic resins may cause contact dermatitis and damage the pulp. Avoid prolonged contact with skin, eyes and soft tissue. Wash thoroughly with water after contact.

STORAGE
Store at ambient temperature.

SHELF LIFE
Refer to outer package for expiration date. Do not use after expiration date.

Wallingford, CT 06492 USA
tel: 203.265.7397

Pentron Clinical

CEpartner4U, 3951DB; 13. NL.
tel: +31 (0)6.516.536.26

REP 13